

Thrombodynamics Analyser



Thrombodynamics assay

12.02.2018 11:43

Measurement title: Hypercoagulation TD
Plasma type: Fresh PFP
Sample code: Example TDX #4

TDX

Fibrin dynamics parameters	Unit	Parameter	Value	Norm.
Rate of clot growth	um/min	V	37,8	20,0-29,0
Lag time	min	Tlag	1,0	0,6-1,5
Initial rate of clot growth	um/min	Vi	61,9	38,0-56,0
Stationary rate of clot growth	um/min	Vst	37,8	20,0-29,0
Clot size at 30 min	um	CS	1440	800-1200
Clot density	a.u.	D	26990	15000-32000
Spontaneous clotting	min	Tsp		30,0-



Comment:

Tech info:

ATTENTION: Research use only. Measurement results shouldn't be used for diagnostic purposes.

S/N: 000000, Operator: Operator, report created 29.03.2019 12:12



48a55d9f-1695-4c13-8415-534d4a6296aa